



# Product Specification Sheet



APS440

Rev. 2 - 08/2003



**IR155B**

## IR155B Single Pull-Cord Resident Station

The Alpha Communications® IR155B Emergency Resident Station provides for efficient use of the nursing staff by permitting fast and reliable communication between nurse and resident.

Calling the nurse is accomplished by pulling the 6-foot-long nylon pull cord mounted on the station. Call-placed assurance and intercom operation is indicated by reliable solid-state LED indicators.

The speaker/microphone is sensitive enough to assure clear conversations even with weak residents. Calls are cancelled (only at the point of origin) by simply sliding the REST switch to the 'LEFT' position.

### FEATURES

- Sensitive speaker/microphone
- Intercom indicator lights when room is being monitored
- Attractive ABS plastic faceplate
- Call-placed LED indicator
- Built-in 6 foot nylon Emergency pull cord.

### SPECIFICATIONS

<b>Dimensions:</b>	Height: 4.50" (119mm) Width: 6.38" (162mm) Depth: 2.75" (70mm) Projects: 0.50" (12.7mm)
<b>Construction:</b>	Attractive, flame retardant type ABS plastic faceplate
<b>Finish:</b>	Light texture, fog white
<b>Speaker:</b>	2.50" x 2.50", 60 ohm with voice response
<b>Connections:</b>	Plug-in type
<b>Housing:</b>	Alpha Communications® IH353 Box with IH352 Ring or Steel City #H3BD with #3GC plaster ring, or EXACT EQUAL. Horizontal mount.

### System Notes

Refer to NC150N and NC200N audio-visual nurse-call systems for more information.

*Due to continuous product improvement, all colors, sizes, materials, finishes and specifications are subject to change without notice.*

### ARCHITECTS' AND ENGINEERS' SPECIFICATIONS

Emergency Resident Station(s) shall be Alpha Communications®/Tektone model IR155B, or approved equal. Model IR155B shall include solid-state call-placed and intercom in-use indicators a 6-foot-long nylon pull cord for call placement, and a 2.50" speaker/microphone with voice response.

All station components shall be mounted on an attractive flame retardant type fog white ABS plastic faceplate. Connections shall be plug-in be plug-in type.

The station shall utilize a 2.50" x 2.50" speaker/microphone for accurate voice reproduction.

Contractor shall observe all local and national electrical and building codes.

The IR155B shall be UL® Listed to the latest 4th Edition Standards.